

CAUTION

Normal Oxygen Level: 20.9%

Low Level Alarm: 19.5%

High Level Alarm: > 23.5%

In all cases, assume the alarm to be valid

Do NOT attempt to repair a faulty sensor.
Report it to EH&S

If Oxygen Sensor is in Alarm:

Call Public Safety: (212) 305 - 7979

During Normal Business Hours, also call EH&S: (212) 305-6780

In Room Alarm Activation	Alarm Activated Prior to Entry
Evacuate workers in room and close the door	DO NOT enter the room
Be prepared to provide information to responders	Be prepared to provide information to responders

PI: _____

Tel #: _____

Quick Dial

Public Safety >



< EH&S

Liquid Nitrogen
FAQ Sheet >



Have more questions? Reach out to EH&S at Labsafety@columbia.edu or give us a call